



Product Service

# CERTIFICATE

No. Z1 17 01 83766 005

**Holder of Certificate:** **Biobase Biodustry (Shandong) Co., Ltd.**  
 No. 2717 Gongyeyi Road, Mingshui Economic Developing Zone  
 250200 Zhangqiu, Jinan, Shandong  
 PEOPLE'S REPUBLIC OF CHINA

**Production Facility(ies):** 83766

**Certification Mark:**



**Product:** **Laboratory Equipment  
 Microbiological Safety Cabinet**

**Model(s):** **BSC-3FA2, BSC-4FA2**

**Parameters:**

Rated input voltage:	230VAC
Rated input frequency:	50Hz
Rated input power:	1050W (BSC-3FA2) 1150W (BSC-4FA2)
Class of MSC:	II
Electrical protection class:	I

**Tested according to:** EN 12469:2000  
 EN 61010-1:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 884001541801

**Valid until:** 2022-01-05

**Date,** 2017-01-12

( Feipeng Lin )

Page 1 of 1



04052767261664





China

TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
M Building, No. 7 Wangjing Zhonghuan Nanlu, Chaoyang District, Beijing 100102, P.R. China

Choose certainty.  
Add value.

## Biobase Biodustry (Shandong) Co., Ltd.

Your reference/letter of	Our reference/name	Tel. extension/e-mail	Fax extension	Date	Page
884001541801	Mr. Lin Feipeng	+86 10 65906186 Feipeng.Lin@tuv-sud.cn	+861065906182	13 January 2017	1/1

### **Subject: Your TÜV SÜD certificate**

Dear Sir/Madam,

We are pleased to send you your below listed certificate(s), which entitle you to label your certified product(s) with the respective certification mark as applicable.

<b><u>Certificate No.</u></b>	<b><u>Product</u></b>	<b><u>Model</u></b>
Z1.17.01.83766.005	Microbiological Safety Cabinet	BSC-3FA2, BSC-4FA2

Further, we like to remind you on following information, which are applicable to all certificate holders:

- As a certificate holder, you have contractual obligations to the related TÜV SÜD certification body as outlined in global "Testing and Certification Regulations" of TÜV SÜD, which you can receive from us or download at: <http://www.tuv-sud.cn/cn-en/terms-and-conditions>
- If the products listed on the certificate will be exported to European Market, the name of the certificate holder and the name and address of the importer or authorized representative based within the European Economic Area must be clearly printed on the product and in the user manual, or where that is not possible, on the packaging and user manual.
- To ensure that the validity of the certificate is not jeopardized, please inform us about any change in your production or of the product itself (this also includes changes to the user manual or rating label)
- If you want to stop the renewal of the above listed certificate(s) for the following year, please inform TÜV SÜD in writing by November 15 of the current year. If not received on time, the annual fee of the following year is to be paid and the cancellation will be done the year after.

If you have any question regarding your certificate(s) or the proper use of the certification mark on your products, please do not hesitate to contact us.

Yours sincerely,  
TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
Su Qiao, Assistant  
ENE & IDP

GCN\_TFS\_F\_00.01 - Rev. 0

HSBC Bank (China) Company Limited, Beijing Branch  
Account No. (RMB): 626-228472-011  
Account No. (USD): 006-228472-055  
Account No. (EUR): 006-228472-056  
SWIFT Code: HSBCCN5HBJG

Managing Director of  
North Region:  
Mr. Zhu Wencai

Telephone: +86 10 6590 6186  
Telefax: +86 10 6590 6182  
[www.tuv-sud.cn](http://www.tuv-sud.cn)



TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
TÜV SÜD Group

M Building, No. 7 Wangjing Zhonghuan Nanlu  
Chaoyang District, Beijing 100102, P.R. China



Product Service

# Attestation of Conformity

No. E8A 16 12 83766 003

**Holder of Certificate:** **Biobase Biodustry (Shandong) Co., Ltd.**  
 No. 2717 Gongyeyi Road, Mingshui Economic Developing Zone  
 250200 Zhangqiu, Jinan, Shandong  
 PEOPLE'S REPUBLIC OF CHINA

**Name of Object:** **Laboratory Equipment  
 Microbiological Safety Cabinet**

**Model(s):** **BSC-3FA2, BSC-4FA2**


**Description of Object:**


Rated voltage:	230VAC, 50Hz
Rated power input:	1050W (BSC-3FA2) 1150W (BSC-4FA2)
Protection class:	I

**Tested according to:** EN 61326-1:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

**Test report no.:** 888001541901





**Date,** 2016-12-27 (Jing Zhang)

**CE** After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. That declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1



Greater China

TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
M Building, No. 7 Wangjing Zhonghuan Nanlu, Chaoyang District, Beijing 100102, P.R. China

**Choose certainty.  
Add value.**

## Biobase Biodustry (Shandong) Co., Ltd.

Your reference/letter of	Our reference/name	Tel. extension/e-mail	Fax extension	Date	Page
888001541901	Ms. Zhang Jing	+861065906186 Jing.Zhang@tuv-sud.cn	+861065906182	27 December 2016	1/1

### **Subject: Your Attestation of conformity**

Dear Sir/Madam,  
We are pleased to send you your below listed certificate(s) of conformity.

<b><u>Certificate No.</u></b>	<b><u>Product</u></b>	<b><u>Model</u></b>
E8A.16.12.83766.003	Microbiological Safety Cabinet	BSC-3FA2, BSC-4FA2

- Further, we like to remind you on following information, which are applicable to all certificate holders:
- As a certificate holder, you have contractual obligations to the related TÜV SÜD certification body as outlined in global "Testing and Certification Regulations" of TÜV SÜD, which you can receive from us or download at: <http://www.tuv-sud.cn/cn-en/terms-and-conditions>
  - If the products listed on the certificate will be exported to European Market, the name of the certificate holder and the name and address of the importer or authorized representative based within the European Economic Area must be clearly printed on the product and in the user manual, or where that is not possible, on the packaging and user manual.

If you have any question regarding your certificate(s) or the proper use of the certification mark on your products, please do not hesitate to contact us.

Yours sincerely,  
TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
Su Qiao  
Assistant  
ENE & IDP

GCM\_TPS\_F\_0001 - Rev. 0

HSBC Bank (China) Company Limited, Beijing Branch  
Account No. (RMB): 626-228472-011  
Account No. (USD): 006-228472-055  
Account No. (EUR): 006-228472-056  
SWIFT Code: HSBCCN5HBJG

Managing Director of  
North Region:  
Mr. Zhu Wencai

Telephone: +86 10 6590 6186  
Telefax: +86 10 6590 6182  
[www.tuv-sud.cn](http://www.tuv-sud.cn)



TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
TÜV SÜD Group

M Building, No. 7 Wangjing Zhonghuan Nanlu  
Chaoyang District, Beijing 100102  
P.R. China



Product Service

# Attestation of Conformity

No. N8A 17 01 83766 004

**Holder of Certificate:** **Biobase Biodustry (Shandong) Co., Ltd.**  
No. 2717 Gongyeyi Road, Mingshui Economic Developing Zone  
250200 Zhangqiu, Jinan, Shandong  
PEOPLE'S REPUBLIC OF CHINA

**Product:** **Laboratory Equipment  
Microbiological Safety Cabinet**

**Model(s):** **BSC-3FA2, BSC-4FA2**

**Parameters:**

Rated input voltage:	230VAC
Rated input frequency:	50Hz
Rated input power:	1050W (BSC-3FA2) 1150W (BSC-4FA2)
Class of MSC:	II
Electrical protection class:	I

**Tested according to:** EN 12469:2000  
EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

**Test report no.:** 884001541801



**Date,** 2017-01-12

( Feipeng Lin )



After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1



China

TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
M Building, No. 7 Wangjing Zhonghuan Nanlu, Chaoyang District, Beijing 100102, P.R. China

Choose certainty.  
Add value.

## Biobase Biodustry (Shandong) Co., Ltd.

Your reference/letter of	Our reference/name	Tel. extension/e-mail	Fax extension	Date	Page
884001541801	Mr. Lin Feipeng	+861065906186 Feipeng.Lin@tuv-sud.cn	+861065906182	13 January 2017	1/1

### Subject: Your Attestation of conformity

Dear Sir/Madam,

We are pleased to send you your below listed certificate(s) of conformity.

<u>Certificate No.</u>	<u>Product</u>	<u>Model</u>
N8A.17.01.83766.004	Microbiological Safety Cabinet	BSC-3FA2, BSC-4FA2

- Further, we like to remind you on following information, which are applicable to all certificate holders:
- As a certificate holder, you have contractual obligations to the related TÜV SÜD certification body as outlined in global "Testing and Certification Regulations" of TÜV SÜD, which you can receive from us or download at: <http://www.tuv-sud.cn/cn-en/terms-and-conditions>
  - If the products listed on the certificate will be exported to European Market, the name of the certificate holder and the name and address of the importer or authorized representative based within the European Economic Area must be clearly printed on the product and in the user manual, or where that is not possible, on the packaging and user manual.

If you have any question regarding your certificate(s) or the proper use of the certification mark on your products, please do not hesitate to contact us.

Yours sincerely,  
TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
Su Qiao  
Assistant  
ENE & IDP

GCN\_TPS\_F\_0001 - Rev. 0

HSBC Bank (China) Company Limited, Beijing Branch  
Account No. (RMB): 626-228472-011  
Account No. (USD): 006-228472-055  
Account No. (EUR): 006-228472-056  
SWIFT Code: HSBCCNSHBJG

Managing Director of  
North Region:  
Mr. Zhu Wencai

Telephone: +86 10 6590 6186  
Telefax: +86 10 6590 6182  
[www.tuv-sud.cn](http://www.tuv-sud.cn)



TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch  
TÜV SÜD Group

M Building, No. 7 Wangjing Zhonghuan Nanlu  
Chaoyang District, Beijing 100102, P.R. China